





REV.	REVISION NOTE	DATE	CHECKED
B	Add P/N	12/05/90	



B-B 10 : 3




(X4: 2 Disks/Chamber)(X41: 4 Disks/Chamber)



P/N.:	PKG		Height	Test	
	X.	Y.	H.	P1.	P2.
X4(1)-252008-1612	2.50	2.00	0.80	1.65	1.25
X4(1)-2520085-1612	2.50	2.00	0.85	1.65	1.25
X4(1)-252007-215130	2.50	2.00	0.70	2.15	1.30
X4(1)-252008-207125	2.50	2.00	0.80	2.07	1.25
X4(1)-252009-1612	2.50	2.00	0.90	1.65	1.25
X4(1)-322509-2518	3.20	2.50	0.90	2.50	1.80
X4(1)-322509-2116	3.20	2.50	0.90	2.15	1.60
X4(1)-322509-2516	3.20	2.50	0.90	2.50	1.60
X4(1)-322510-265134	3.20	2.50	1.00	2.65	1.34
X4(1)-322511-2116	3.20	2.50	1.10	2.15	1.60
X4(1)-322516-2617	3.20	2.50	1.60	2.60	1.70
X4(1)-503210-2522	5.00	3.20	1.00	2.54	2.20
X4(1)-503210-3723	5.00	3.20	1.00	3.70	2.30
X4(1)-503211-2522-6P	5.00	3.20	1.10	2.54	2.20
X4(1)-5032115-3723	5.00	3.20	1.15	3.70	2.30
X4(1)-503212-2522	5.00	3.20	1.20	2.54	2.20
X4(1)-705013-5038	7.00	5.00	1.30	5.08	3.80
X4(1)-705014-5038	7.00	5.00	1.40	5.08	3.80
X4(1)-734810-5036	7.30	4.80	1.00	5.08	3.60
X4(1)-755013-5038	7.50	5.00	1.30	5.08	3.80
X4(1)-7551195-5038	7.50	5.10	1.95	5.08	3.80
X42-705013-5038-6P	7.00	5.00	1.30	5.08	3.80

Note:
X4: Oscillator temperature disk pallet for W-2800. (2 Disks/Chamber)
X41: Oscillator temperature disk pallet for W-2800B. (4 Disks/Chamber)
X42 : Oscillator temperature disk pallet for LVDS/LVPECL CXO. (6 positions)
Gold plating on the golden finger >25u".
Temperature range is -45~140 °C.

 XXX. ± 0.5 XX. ± 0.3 X. ± 0.2 .X ± 0.1 .XX ± 0.05 ANGLES ± 1/4" DRILL HOLD +0.15 / -0	TITLE		W-2800 Temperature Test Disk Pallet		A FORCE 優力精密工業有限公司 TEL: 886-3-6571238 FAX: 886-2-81926187
	DWG NO.	X4-XXXXXX-XXXX	REV.	A	
	MATERIAL-		PAPER	A4	
	UNIT: mm	SCALE: 1:3	SHEET: 1 of 1		
	APPROVAL	DATE	DESIGNER	DATE	
	Steven Yang	15/06/22			